

Lactanet - Valacta

555, boul. des Anciens-Combattants
 Sainte-Anne-de-Bellevue, QC H9X 3R4
 T 514-459-3030 ext 7990
 Email: microlab@lactanet.ca
 1-800-bon-lait



ATTN: Lab manager / supervisors
 Date: 2020-08-17

Somatic cell count standards
 Batch: SCC-08-2020



Certificate of analysis for direct microscopic cell count (DMSCC)

Standard	DMSCC results* (cells/ml)
Low (blue)	138 000
Medium (yellow)	451 000
High (red)	844 000

* Results are to be considered as estimates only

For best results:

- Expiry date: 2020-09-15
- Do not subdivide samples into smaller portion
- Warm to $\pm 40^{\circ}\text{C}$ and shake gently to disperse butterfat
- Discard used vials
- Do not re-heat samples
- Do not use leaking vials

<input type="checkbox"/> M-H. Castonguay, M.Sc., Mcb.A.		<div style="text-align: right;">  2020-08-17 </div>
<input checked="" type="checkbox"/> K. M. Amine, M.Sc., Mcb.A.		

Cette version remplace et annule toute version antérieure, le cas échéant / If applicable, this version replaces all previous issues.

Ce document est confidentiel / This document is confidential.

Les résultats et commentaires, le cas échéant, ne se rapportent qu'aux échantillons décrits plus haut / Results and comments, if applicable, are only related to the above samples.

Ce certificat ne doit pas être reproduit, sinon en entier, sans l'autorisation écrite du laboratoire / This certificate shall not be reproduced except in full, without written approval of the laboratory.

Fin du rapport / End of report