

**Lactanet - Valacta**

555, boul. des Anciens-Combattants  
Sainte-Anne-de-Bellevue, QC H9X 3R4  
T 514-459-3030 ext 7990  
Email: microlab@lactanet.ca  
1-800-bon-lait





ATTN: Lab manager / supervisors  
Date: 2021-03-15

Instrumental counts of canadian fluoro-opto-electronic counters  
Batch : SCC-03-2021

Standard	Results (count/ml)		
	lowest	average	highest
Low (blue)	114 000	137 000	157 000
Medium (yellow)	377 000	466 000	521 000
High (red)	584 000	794 000	851 000

For best results:

- Expiry date: 2021-04-20
- Do not subdivide samples into smaller portion
- Warm to ± 40°C and shake gently to disperse butterfat
- Discard used vials
- Do not re-heat samples
- Do not use leaking vials

		
<input checked="" type="checkbox"/> M-H. Castonguay, M.Sc., Mcb.A.	<input type="checkbox"/> K. M. Amine, M.Sc., Mcb.A.	

Cette version remplace et annule toute version antérieure, le cas échéant / If applicable, this version replaces all previous issues.

Ce document est confidentiel / This document is confidential.

Les résultats et commentaires, le cas échéant, ne se rapportent qu'aux échantillons décrits plus haut / Results and comments, if applicable, are only related to the above samples.

Ce certificat ne doit pas être reproduit, sinon en entier, sans l'autorisation écrite du laboratoire / This certificate shall not be reproduced except in full, without written approval of the laboratory.

Fin du rapport / End of report