

**Lactanet - Valacta**

555, boul. des Anciens-Combattants  
 Sainte-Anne-de-Bellevue, QC H9X 3R4  
 T 514-459-3030 ext 7990  
 Email: microlab@lactanet.ca  
 1-800-bon-lait



ATTN: Lab manager / supervisors  
 Date: 2021-04-19

Somatic cell count standards  
 Batch: SCC-04-2021

Certificate of analysis for direct microscopic cell count (DMSCC)

Standard	DMSCC results* (cells/ml)
Low (blue)	142 000
Medium (yellow)	487 000
High (red)	893 000

\* Results are to be considered as estimates only

For best results:

- Expiry date: 2021-05-25
- Do not subdivide samples into smaller portion
- Warm to  $\pm 40^{\circ}\text{C}$  and shake gently to disperse butterfat
- Discard used vials
- Do not re-heat samples
- Do not use leaking vials



M-H. Castonguay, M.Sc., Mcb.A.

K. M. Amine, M.Sc., Mcb.A.

2021-04-19

Cette version remplace et annule toute version antérieure, le cas échéant / If applicable, this version replaces all previous issues.

Ce document est confidentiel / This document is confidential.

Les résultats et commentaires, le cas échéant, ne se rapportent qu'aux échantillons décrits plus haut / Results and comments, if applicable, are only related to the above samples.

Ce certificat ne doit pas être reproduit, sinon en entier, sans l'autorisation écrite du laboratoire / This certificate shall not be reproduced except in full, without written approval of the laboratory.

Fin du rapport / End of report