



Valacta

555, boul. Des Anciens-Combattants
Sainte-Anne-de-Bellevue, Qc H9X 3R4
T 514-459-3030 #7990 Email : mhcastonguay@valacta.com
F 514-459-3023
1-800-bon-lait

ATTN: Lab manager / Supervisor

Date: February 19, 2019

**Somatic cell count standards
Batch # SCC-02-2019 shipped to you on February 13, 2019.**

Certificate of analysis

Range	Batch #	DMSCC results*
Low - blue	SCC-02-2019	155,000 cells/ml
Medium - yellow	SCC-02-2019	469,000 cells/ml
High - red	SCC-02-2019	928,000 cells/ml

*Results are to be considered as estimates only

For best results:

- Expiry date: March 19, 2019.
- Do not sub-divide samples into smaller portion
- Warm to $\pm 40^{\circ}\text{C}$ and shake gently to disperse butterfat
- Discard used vials
- Do not re-heat samples
- Do not use leaking vials

Unless we are notified before March 7, 2019; a new set will be sent to your lab on March 13, 2019.

Marie-Hélène Castonguay, M.Sc. Microbiologist

