



Valacta

555, boul. Des Anciens-Combattants
Sainte-Anne-de-Bellevue, Qc H9X 3R4
T 514-459-3030 #7990 Email : mhcastonguay@valacta.com
F 514-459-3023
1-800-bon-lait

ATTN: Lab manager / Supervisor

Date: March 18, 2019

**Somatic cell count standards
Batch # SCC-03-2019 shipped to you on march 13, 2019.**

Certificate of analysis

Range	Batch #	DMSCC results*
Low - blue	SCC-03-2019	142,000 cells/ml
Medium - yellow	SCC-03-2019	456,000 cells/ml
High - red	SCC-03-2019	901,000 cells/ml

*Results are to be considered as estimates only

For best results:

- Expiry date: April 16, 2019.
- Do not sub-divide samples into smaller portion
- Warm to $\pm 40^{\circ}\text{C}$ and shake gently to disperse butterfat
- Discard used vials
- Do not re-heat samples
- Do not use leaking vials

Unless we are notified before April 4, 2019; a new set will be sent to your lab on April 10, 2019.



Khalie Mahamad Amine, M.Sc. Microbiologist