



# Valacta

555, boul. Des Anciens-Combattants  
Sainte-Anne-de-Bellevue, Qc H9X 3R4  
T 514-459-3030 #7990 Email : mhcastonguay@valacta.com  
F 514-459-3023  
1-800-bon-lait

---

**ATTN: Lab manager / Supervisor**

**Date: February 19, 2019**

**Instrumental counts<sup>a</sup> of somatic cells of calibration standards  
Batch # SCC-02-2019 shipped to you on Wednesday February 13, 2019.**

Batch #	DMSCC results	Compared instrumental results <sup>b</sup>		
		Lowest	Average	Highest
Low # SCC-02-2019 (blue)	<b>155</b>	149	<b>170</b>	188
Medium # SCC-02-2019 (yellow)	<b>469</b>	443	<b>485</b>	549
High # SCC-02-2019 (red)	<b>928</b>	874	<b>947</b>	1033

<sup>a</sup> For final count: x 1000 cells/ml

<sup>b</sup> All instrumental values suppose the use of the current calibration and not yet re-calibrated.

Number of instruments: 13

No outlier detected.

We thank all the laboratories for offering their instrumental results.

Marie-Hélène Castonguay, M.Sc. Microbiologist

