



Valacta

555, boul. Des Anciens-Combattants
Sainte-Anne-de-Bellevue, Qc H9X 3R4
T 514-459-3030 #7990 Email : mhcastonguay@valacta.com
F 514-459-3023
1-800-bon-lait

ATTN: Lab manager / Supervisor

Date: March 18, 2019

**Instrumental counts^a of somatic cells of calibration standards
Batch # SCC-03-2019 shipped to you on Wednesday march 13, 2019.**

Batch #	DMSCC results	Compared instrumental results ^b		
		Lowest	Average	Highest
Low # SCC-03-2019 (blue)	142	135	159	183
Medium # SCC-03-2019 (yellow)	456	475	509	550
High # SCC-03-2019 (red)	901	855	915	978

^a For final count: x 1000 cells/ml

^b All instrumental values suppose the use of the current calibration and not yet re-calibrated.

Number of instruments: 15

No outlier detected.

We thank all the laboratories for offering their instrumental results.



Khalie Mahamad Amine, M.Sc. Microbiologist