



# Valacta

555, boul. Des Anciens-Combattants  
Sainte-Anne-de-Bellevue, Qc H9X 3R4  
T 514-459-3030 #7990 Email : mhcastonguay@valacta.com  
F 514-459-3023  
1-800-bon-lait

**ATTN: Lab manager / Supervisor**

**Date: May 20, 2019**

**Somatic cell count standards  
Batch # SCC-05-2019 shipped to you on May 15, 2019.**

### Certificate of analysis

Range	Batch #	DMSCC results*
Low - blue	SCC-05-2019	120,000 cells/ml
Medium - yellow	SCC-05-2019	445,000 cells/ml
High - red	SCC-05-2019	884,000 cells/ml

\*Results are to be considered as estimates only

#### For best results:

- Expiry date: June 18, 2019.
- Do not sub-divide samples into smaller portion
- Warm to  $\pm 40^{\circ}\text{C}$  and shake gently to disperse butterfat
- Discard used vials
- Do not re-heat samples
- Do not use leaking vials

Unless we are notified before June 6, 2019; a new set will be sent to your lab on June 12, 2019.

Marie-Hélène Castonguay, M.Sc. Microbiologist

