



# Valacta

555, boul. Des Anciens-Combattants  
Sainte-Anne-de-Bellevue, Qc H9X 3R4  
T 514-459-3030 #7990 Email : mhcastonguay@valacta.com  
F 514-459-3023  
1-800-bon-lait

**ATTN: Lab manager / Supervisor**

**Date: May 20, 2019**

**Instrumental counts<sup>a</sup> of somatic cells of calibration standards  
Batch # SCC-05-2019 shipped to you on Wednesday May 15, 2019.**

Batch #	DMSCC results	Compared instrumental results <sup>b</sup>		
		Lowest	Average	Highest
Low # SCC-05-2019 (blue)	<b>120</b>	113	<b>126</b>	159
Medium # SCC-05-2019 (yellow)	<b>445</b>	425	<b>489</b>	527
High # SCC-05-2019 (red)	<b>884</b>	845	<b>909</b>	972

<sup>a</sup> For final count: x 1000 cells/ml

<sup>b</sup> All instrumental values suppose the use of the current calibration and not yet re-calibrated.

Number of instruments: 16

Two outliers detected on two calibration level with no impact on final results.

We thank all the laboratories for offering their instrumental results.

Marie-Hélène Castonguay, M.Sc. Microbiologist

